

arkray

Urinalysis reagent for machine reading

Uriflet S

Uriflet S 11UA



Uriflet S 11UA is the urinalysis reagent for machine reading which can measure 11 items. The urine albumin/creatinine ratio and protein/creatinine ratio can also be calculated.

By the use of UriFlet S 11UA, screening tests can be conducted for ailments such as diabetic nephropathy, including Chronic Kidney Disease (CKD).

Related products

Fully Automated Urine Analyzer

AUTION MAX AX-4060



Solution to Your Needs

AX-4060 shall support simple and fast testing in hospitals, and contribute to the early detection and treatment of kidney ailments such as diabetic nephropathy, including Chronic Kidney Disease (CKD).

Interpretation

Item		Interpretation						
		-	±	1+	2+	3+	4+	
Glucose	Qualitative symbol	-						
	Semiquantitative value (mg/dL)		30	50	70	100	150	200
Albumin	Semiquantitative value (mg/L)	10	30	80	150	OVER		
	Qualitative symbol	-	±	1+	2+	3+	4+	
Protein	Semiquantitative value (mg/dL)		10	20	30	50	70	100
	Qualitative symbol				200	300	600	OVER
Urobilinogen	NORMAL		1+	2+	3+	4+		
	Semiquantitative value (mg/dL)		2	3	4	6	8	12
Bilirubin	Qualitative symbol	-	1+	2+	3+	4+		
	Semiquantitative value (mg/dL)		0.5	1.0	2.0	3.0	4.0	10.0
Creatinine	Semiquantitative value (mg/dL)	10	50	100	200	300	OVER	
pH	Measurement Value	5.0	5.5	6.0	6.5	7.0	7.5	8.0
Blood	Qualitative symbol	-	±	1+	2+	3+		
	Semiquantitative value (mg/dL)		0.03	0.06	0.1	0.2	0.5	1.0
	(Ery/µL)		10	20	60	300		
Ketones	Qualitative symbol	-	±	1+	2+	3+	4+	
	Semiquantitative value (mg/dL)			10	20	40	60	150
Nitrite	Qualitative symbol	-	1+	2+				
Leukocytes	Semiquantitative value (Leu/µL)	-	25	75	250	500		

Protein/Creatinine ratio and Albumin/Creatinine ratio is obtained automatically.

Specification

Name	Instrument urine test strip "UriFlet S 11UA"
Meas. target	Urine
Meas. items	Glu, Pro, Bld, pH, Bil, Uro, Ket, Nit, Leu, Cre, Alb, (Calculation items: Alb/Cre ratio, Pro/Cre ratio)
Packaging unit	100 per pack

ARKRAY GLOBAL BUSINESS, INC.

Yousuien-nai, 59 Gansuin-cho,
Kamigyo-ku, Kyoto 602-0008, JAPAN
TEL 81-75-662-8979 FAX +81-75-431-1202